





HEALTHCARE[®]

Plastic Anti-Ligature Tethers

Designed with patient well-being & safety in mind

CleanGrip Plastic Anti-Ligature Tethers

Safe and easy to grab, the Crest CleanGrip is a highly visible and cleanable alternative to nylon station pullcords. The CleanGrip tether is constructed of 12 mm wide PVC that is 1.8 mm thick. It is easily disinfected by germicidal detergent, disposable wipes or disinfectant cleaner. The anti-ligature safety enhancement is created with strategic placement of "breaks" every 4" along the length of the cord. The weak zones break when 11 pounds of pull-force is exerted minimizing potential safety risks compared to standard pull cords. CleanGrip is available in non-preassembled configurations and in bulk rolls.

- Bulk tether can be cut to any desired length
- Available in red or blue to provide contrast for high visibility
- Adapter connects to CleanCord[™] Pullcords, as well as functioning as a pendant & bead chain adapter

Compare with Sanipull[®] for a better value

Compare with Sanipuli® for a better value CleanGrip Red Tethers & Configurations			LATEX-FREE	1
				YEAR WARRANTY
Part No.	Description	Price		
<u>116687</u>	CleanGrip Anti-Ligature Tether with Spring Clip and Adapter, red, 6'	10.20		-
<u>116686</u>	CleanGrip Anti-Ligature Tether with two Spring Clips, red, 6'	10.48 🗧		-
<u>116602</u>	CleanGrip Anti-Ligature Tether, red, 6', 1 pack (5/pk)	18.68	8	
<u>116387</u>	CleanGrip Anti-Ligature Tether with two Adapters, red, 6'	10.38		
<u>116608</u>	CleanGrip Anti-Ligature Tether with Security Clip and Adapter, red, 6'	10.65		116387
<u>116609</u>	CleanGrip Anti-Ligature Tether with two Security Clips, red, 6'	10.45		
<u>116375</u>	CleanGrip Bulk Anti-Ligature Tether, red, 500'	255.17		
CleanGr	ip Blue Tethers & Configurations	<i>a</i>		-
Part No.	Description	Price		
<u>116675</u>	CleanGrip Anti-Ligature Tether with Spring Clip and Adapter, blue, 6'	10.04		
<u>116674</u>	CleanGrip Anti-Ligature Tether with two Spring Clips, blue, 6'	10.53		116675
<u>116601</u>	CleanGrip Anti-Ligature Tether, blue, 6', 1 pack (5/pk)	18.68		
<u>116158</u>	CleanGrip Anti-Ligature Tether with two Adapters, blue, 6'	10.38		
<u>116606</u>	CleanGrip Anti-Ligature Tether with Security Clip and Adapter, blue, 6'	10.47		
<u>116607</u>	CleanGrip Anti-Ligature Tether with two Security Clips, blue, 6'	10.38		
<u>115840</u>	CleanGrip Anti-Ligature Tether with two Adapters, #6 Bead Chain	13.13		
	and Coupling, blue, 6'			
<u>116157</u>	CleanGrip Bulk Anti-Ligature Tether, blue, 500'	255.17		116676
CleanGr	ip Clips & Adapters			
Part No.	Description	Price		
<u>116676</u>	CleanGrip Spring Clip, white, 1 pack (10/pk)	27.91		
<u>115842</u>	CleanGrip Adapter, attaches to cords or bead chains, red, 1pk (10/pk)	36.25		
<u>116613</u>	CleanGrip Replacement Clip Kit, 1 pack (10/pk)	10.47		
	DECT () 800.328.8908			115842

800.369.9207

CrestHealthcare.com

©2024 Crest Healthcare. All listed prices and products are subject to change without notice. Crest is not responsible for photographic or typographical errors.



Installation Instructions

Video Available on You Tube

#6 Bead Chain

Insert #6 Bead Chain into adapter through small hole on top then attach coupling unit to bead chain. Rest coupling unit within adapter. Fold CleanGrip tether twice. Insert into adapter next to coupling. Snap adapter closed.

Spring Clip

Insert tether through black cover piece and pull through. Take tether and thread through outside slot towards the top of the spring clip. Once through, take tether and thread through inside slot toward bottom of spring clip. **Tip**: **we recommend using a thin nose plier to thread tether through spring clip.** Pull tether to the edge of the imprint and tighten any excess slack while holding threaded end. Slide black cover all the way up and push tight.

Screw & Nut

Trim the CleanGrip tether, if needed, in a straight line. Using a ³/₃₂" drill bit, drill a hole centered ¹/₄" from the end of the CleanGrip tether. Insert provided screw through the security clip and CleanGrip hole. With black sleeve facing downward, secure nut to the screw.

Plastic & Nylon Pullcords

Insert thin nylon cord into adapter through small hole on top. Tie knot and rest cord within adapter to prevent pull through. Fold CleanGrip tether twice. Insert into adapter next to nylon cord. Snap adapter closed.

Tether

Fold CleanGrip tether twice. Insert into adapter and make sure it is snug. Snap adapter closed.



🔇 800.328.8908

🗐 800.369.9207

CrestHealthcare.com