

LI382LED Corridor Zone Light IL898 Specification Sheet

Tek-CARE Nurse Call and CM800 Emergency Call Systems • Rev. 6 - 11/2020

Features

- Long lasting LEDs
- No bulb replacement needed
- Plug-in connector supplied
- Attractive flame-retardant plastic panel

Description

The LI382LED Corridor Zone Light provides visual indication of calls when used with a nurse call system, emergency call system, or similar application.

The LI382LED Corridor Zone Light indicates an emergency call by the rapid flashing of the two red LED bulbs associated with the zone or area from which an emergency call has been placed. Normal calls are indicated by a steady illumination of the two white LEDs. If an emergency and normal call are placed in the same zone at the same time, the red LEDs connected to the emergency station will flash rapidly, while the white LEDs associated with the normal call will maintain a steady illumination.

The standard mounting plate features both one- and twogang mounting options.

Listings

UL[®] 1069 Listed; UL[®] File Number E71599.

Specifications

Dimensions:	4.8 "H \times 5.0"W \times 2.75"D
	$(122 \text{ mm} \times 127 \text{ mm} \times 70 \text{ mm})$
LEDs:	2V 20mA per LED
Power:	24 VDC
Finish:	Base is flame-retardant (UL® 94HB) plastic. Lens
	is flame-retardant translucent polycarbonate
	plastic.
Housing:	Fits standard one- or two-gang box with one- or
	two-gang ring.



LI382LED Corridor Zone Light

Required Components

Tek-CARE[®] NC110, NC150, NC200 or 300II Nurse Call System

Tek-CARE[®]500 Wireless Nurse Call System with NC50xES-series remote receiver CM800 Emergency Call System with CM800-series speaker panel

www.tektone.com

324 Industrial Park Road • Franklin, NC 28734 • tektone@tektone.com Phone: 828.524.9967 • Fax: 828.524.9968 • Sales: Choose option 2 • Tech Support: Choose option 3 TekTone's quality system is registered by DQS to the ISO 9001:2015 standard.